

UC Irvine

Western Journal of Emergency Medicine: Integrating Emergency Care with Population Health

Title

Modification of Daly's Do-it-yourself, Ultrasound-guided Pericardiocentesis Model for Added External Realism

Permalink

<https://escholarship.org/uc/item/0676t4fp>

Journal

Western Journal of Emergency Medicine: Integrating Emergency Care with Population Health, 19(3)

ISSN

1936-900X

Authors

Young, Timothy P.
Kuntz, Heather M.

Publication Date

2018-01-01

DOI

10.5811/westjem.2018.2.37736

License

[CC BY 4.0](#)

Peer reviewed

Modification of Daly's Do-it-yourself, Ultrasound-guided Pericardiocentesis Model for Added External Realism

Timothy P. Young, MD
Heather M. Kuntz, MD

Loma Linda University Medical Center and Children's Hospital, Department of
Emergency Medicine, Loma Linda, California

Section Editor: Mark I. Langdorf, MD, MHPE

Submission history: Submitted January 26, 2018; Accepted February 19, 2018

Electronically published March 13, 2018

Full text available through open access at http://escholarship.org/uc/uciem_westjem

DOI: 10.5811/westjem.2018.2.37736

[West J Emerg Med. 2018;19(3)465–466.]

Comment on:

Daly R, Planas JH, Edens MA. Adapting gel wax into an ultrasound-guided pericardiocentesis model at low cost. *West J Emerg Med.* 2017;18(1):114–6.

We were happy to discover the article by Daly and colleagues describing a low-cost, pericardiocentesis training model (Figure 1).¹ We have struggled to find a cost-effective means of demonstrating and practicing ultrasound-guided pericardiocentesis with our emergency medicine residents. We greatly appreciated their ingenuity in building upon and improving previous do-it-yourself models.^{2,3} We were especially impressed with their addition of a plastic Halloween skeleton thorax and 250 mL normal saline bag to act as the pericardial sac. In that same spirit, we have devised a modification of their model that offers the benefit of more-realistic external landmarks.

We followed their instructions with several exceptions. Instead of the square-shaped plastic container that they used for

a chest wall mold, we used a plastic, manikin dress-form torso (Amazon, \$25; <https://tinyurl.com/amazondressform>). We cut the back of the dress form out with tin snips and turned it upside down to act as a mold. We also used ballistics gel (Amazon, \$60; <https://tinyurl.com/amazonballisticsgel>) instead of gel wax. Ballistics gel does not require the addition of a substance such as flour to simulate the echogenicity of human tissue. Ballistics gel is clear, so we added an optional flesh-colored dye (Humimic, \$30; <https://humimic.com/product-category/dye/>). Like wax gel, ballistics gel can be removed and re-melted to create a new model after repeated needle aspirations. We melted the ballistics gel in a Hamilton-Beach 7-quart cooker (Amazon, \$30; <https://tinyurl.com/amazonHBcooker>) for several hours on the “high” setting, then stirred in the dye and poured the ballistics gel into the turned-over torso (Figure 2). Unlike the Daly model, we did not need to add flour or strain off foam. We placed ice underneath the torso to prevent melting or deforming of the hard plastic torso shell, but this may have been unnecessary.



Figure 1. Original ultrasound-guided pericardiocentesis model from Daly R. et al. *West J Emerg Med.* 2017;18(1):114–6.



Figure 2. Modified ultrasound-guided pericardiocentesis trainer using a plastic dress form torso turned upside down to act as a mold.

Like the Daly model, we were able to build our model for less than \$200. The additional benefit we found was that our model created a realistic chest surface and external landmarks. Our residents were able to practice ultrasound-guided sub-xiphoid and parasternal approaches as well as the blind, landmark-based pericardiocentesis technique. Training programs planning to utilize Daly's brilliant innovation may wish to follow our lead to further increase realism.

Address for Correspondence: Tim Young, MD, Loma Linda University Medical Center and Children's Hospital, Department of Emergency Medicine, 11234 Anderson Street, A108, Loma Linda, CA 92354. Email: tpyoung@llu.edu.

Conflicts of Interest: By the *WestJEM* article submission agreement, all authors are required to disclose all affiliations, funding sources and financial or management relationships that could be perceived as potential sources of bias. No author has professional or financial relationships with any companies that are relevant to this study. There are no conflicts of interest or sources of funding to declare.

Copyright: © 2018 Young et al. This is an open access article distributed in accordance with the terms of the Creative Commons Attribution (CC BY 4.0) License. See: <http://creativecommons.org/licenses/by/4.0/>

REFERENCES

1. Daly R, Planas JH, Edens MA. Adapting gel wax into an ultrasound-guided pericardiocentesis model at low cost. *West J Emerg Med.* 2017;18(1):114-6.
2. Zerth H, Harwood R, Tommaso L, et al. An inexpensive, easily constructed, reusable task trainer for simulating ultrasound-guided pericardiocentesis. *J Emerg Med.* 2012;43(6):1066-9.
3. Campo Dell'orto M, Hempel D, Starzetz A, et al. Assessment of a low-cost ultrasound pericardiocentesis model. *Emerg Med Int.* 2013;2013:376415.